

Guidance on Adoption of ISO 80369-3 Standard ENFit® Connectors in California

To reduce the risk of wrong route delivery of fluids and gases (tubing misconnections) there is an ongoing effort led by the International Organization for Standardization (ISO) to address small-bore connectors for healthcare applications. The objective of the ISO 80369 series of standards is to specify designs of small-bore connectors for various clinical applications to reduce the likelihood of tubing misconnections. The Final Draft International Standard 80369-3 has been approved and published. The standard provides guidance to manufacturers for production of an enteral connector that meets California legislation AB444.

Effective immediately, as stated in AB 444¹, “Health facilities including general acute care hospitals, acute psychiatric hospitals, skilled nursing facilities, and special hospitals in the State of California² will be prohibited to use enteral feeding devices with connectors which would be compatible to connect with devices intended for applications other than enteral”. The ISO 80369-3 standard connector design in the female to male orientation that addresses this legislation is a registered trademark name commonly known as ENFit.

ENFit Tip Syringes are critical to support the introduction of feeding tubes with ENFit connectors. Syringes are used for flushing, hydration, bolus feeding and enteral administration of medication. Feeding tubes and medication ports on feeding sets with ENFit male connectors will not be compatible with current Luer, oral, or catheter tip syringes. While one large syringe supplier has indicated they will not support this ENFit worldwide patient safety initiative³, many other suppliers have been aggressively working to provide adequate supply of ENFit Tip Syringes to collectively meet market demand.

Recent concerns have been raised regarding accuracy of low doses of medication (< 2.0mL). A technical team of industry experts have worked collaboratively to identify a solution to address dose accuracy with a new ENFit Low Dose Tip Syringe design. This design has been validated and demonstrated the ability to deliver an accurate dose consistent with current practice with 95% confidence, fits appropriately into current practice and maintains compatibility with other ENFit devices. The LDT syringe has been reviewed and gained FDA 510(k) clearance for two manufacturers and is now ready for market introduction and support the broader transition to ENFit.

The conversion to the ENFit connector system impacts the entire enteral feeding system across all health care settings. To avoid disruption of therapy, a careful and methodical transition to new connectors is recommended over the course of 2016 across the United States. Introduction of ENFit may vary depending on your supplier(s) timing. GEDSA encourages manufacturers to introduce, and healthcare facilities to adopt ENFit Tip Syringes and feeding tubes as soon as possible to meet the California mandate. GEDSA and its supporting organizations strongly suggest you work with your supplier representative and distributor network to understand their specific plans for conversion. In particular, you must confirm that your syringe supplier has adequate supply of syringes before you convert to ENFit feeding tubes.

¹ CA Assembly Bill 444, 2015 Cal.Stat. 2015

² California Health & Safety Code, 1279.7, section 1250

³ Becton Dickinson Enteral Syringes: Updated Letter to Customers, October 2015

<http://www.bd.com/hypodermic/products/enteral/pdfs/customer-letter-October-2015.pdf>

The Global Enteral Device Supplier Association (GEDSA) is a 501(c)6 nonprofit trade association formed to help introduce international standards for healthcare tubing connectors. Comprised of manufacturers, distributors, and suppliers of enteral nutrition devices worldwide, GEDSA facilitates information flow about the three-phase initiative, which is designed to increase patient safety and optimal delivery of enteral feeding by reducing the risk of tubing misconnections.

ENFit is a federally registered trademark of GEDSA in multiple jurisdictions throughout the world.

GEDSA Members:

A. Hopf GmbH	Cook Medical	Moog
Abbott Nutrition	Dale Medical	NeoMed
Alcor Scientific	Degania Medical	Nestle Health Science
Amsino	Enteral UK	Nutricia
B Braun	Fresenius-Kabi	Qosina
Bard	Halyard Health	Smith's Medical
Baxter	Intervene	UComfor
Boston Scientific	Medela	Vesco Medical
Cair Lgl	Medicina	VR Medical/Ameritus
Cedic	Medline	Vygon
Codan	Medtronic	Xeridiam
Corpak MedSystems		

Supporting Organizations:

AAMI	CHPSO	NNNG
AHRMM	EPSC	NPSF
A.S.P.E.N.	Feeding Tube Awareness Foundation	Oley Foundation
ASHP	HealthTrust	PENG
ASHRM	ISMP	PINNT
AVA	The Joint Commission	Premier
BAPEN	MNI	Vizient
CHA	NHS	